



Contribution ID: 1032

Type: Oral (Non-Student) / orale (non-étudiant)

## The Accurate Determination of Radionuclides without Prior Chemical Separation of Interferences using an Agilent 8800 ICP-MS/MS

Wednesday 15 June 2016 09:00 (15 minutes)

This paper summarizes the accomplishments-to-date of the Agilent 8800 ICP-MS/MS in the direct determination of the radionuclides,  $^{90}\text{Sr}^+$ ,  $^{129}\text{I}^+$ ,  $^{137}\text{Cs}^+$ ,  $^{238}\text{Pu}^+$ ,  $^{239}\text{Pu}^+$ , and  $^{236}\text{U}^+/\text{}^{238}\text{U}^+$  ratios by tandem mass spectrometry with chemical resolution using gas-phase ion chemistry.

The ICP-MS/MS configuration consists of a collision reaction cell (CRC) between two quadrupole mass filters, Q1 and Q2. The abundance sensitivity afforded by tandem MS is better than  $10\text{E}-10$ , eliminating wing overlaps of  $^{127}\text{IH}^+$  on  $^{129}\text{I}^+$  and  $^{238}\text{U}^+$  on  $^{237}\text{Np}^+$  and  $^{239}\text{Pu}^+$ .

The CRC is filled with a gas that reacts with the analyte and its interferences at different rates. When the reaction kinetics favor the analyte, the analyte is converted to a molecular ion which is then measured in "mass-shift" mode. The interference from  $^{235}\text{UH}^+$  on  $^{236}\text{U}^+$  was circumvented by oxidizing U with  $\text{O}_2$ . The  $^{236}\text{U}^+/\text{}^{238}\text{U}^+$  ratio was determined as a  $^{236}\text{U}^{16}\text{O}^+/\text{}^{238}\text{U}^{16}\text{O}^+$  ratio.

When the reaction kinetics favor the interferent, the interferent is converted to another form and the analyte is measured at its elemental mass (the "on-mass" mode). Differences in the oxidizing efficiencies of  $\text{N}_2\text{O}$ ,  $\text{O}_2$ ,  $\text{CO}_2$  were used to discriminate interferences  $\text{Ba}^+$  from  $\text{Cs}^+$ ,  $^{90}\text{Zr}^+$  from  $^{90}\text{Sr}^+$ , and  $^{239}\text{U}^+$  from  $^{239}\text{Pu}^+$ , respectively. The  $^{238}\text{U}^+$  interference on  $^{238}\text{Pu}^+$  was removed by reacting  $\text{U}^+$  to amine cluster ions by a blend of 10%  $\text{NH}_3/90\%$  He. The  $^{129}\text{Xe}^+$  interference on  $^{129}\text{I}^+$  was removed by charge transfer reaction with  $\text{O}_2$ .

Q1 is unique to the Agilent 8800x ICP-MS/MS: it is a unit mass resolution mass spectrometer, operated under vacuum, and precedes the reaction cell. This configuration is vital for the successful and simple implementation of chemical resolution in ICP-MS. Results from the efforts of many scientists will be presented.

**Author:** Ms WEE, Pamela (Agilent Technologies Canada Inc)

**Presenter:** Ms WEE, Pamela (Agilent Technologies Canada Inc)

**Session Classification:** W1-6 Instrumentation for the Detection of Low-Level Radioactivity (DIMP) / Appareillage de détection de radioactivité de faible intensité (DPIM)

**Track Classification:** Instrumentation and Measurement Physics / Physique des instruments et mesures (DIMP-DPIM)