



Canadian Association
of Physicists

Association canadienne
des physiciens et physiciennes

Contribution ID: 2500 Type: **Oral Competition (Graduate Student) / Compétition orale (Étudiant(e) du 2e ou 3e cycle)**

A Supernova Calibration Source for SNO+

Monday 3 June 2019 13:45 (15 minutes)

Only one supernova neutrino burst has ever been detected, and the detection of additional neutrinos from galactic core-collapse supernovae are expected to provide insight on the supernova explosion mechanism. One candidate for detecting supernova neutrinos is SNO+, a multipurpose ultra-low background particle detector. Within SNO+, a galactic supernova neutrino burst is expected to generate an unprecedented rate. Thus, it is necessary to stress-test and optimize the SNO+ data acquisition and electronics so that a supernova signal can be reliably read out. For this purpose, a Supernova Calibration Source is under development to mimic the light expected from supernova neutrino interactions [1]. Using one-dimensional simulated supernova neutrino datasets [2, 3], light profiles representing neutrino interactions are calculated and realised using a laser diode light source delivered into the detector via fibre optics and a deployed light diffuser. This talk will focus on the software conversion of neutrino datasets to light profiles, which define the light intensity and timing in the calibration source.

[1] C. Darrach. The SNO+ Supernova Calibration Source: Development and Testing. MSc Thesis, Laurentian University, Sudbury, ON, Canada (2016)

[2] L. Hübepohl, B. Müller, H.-T. Janka, A. Marek, and G. G. Raffelt. Phys. Rev. Lett. 104, 251101 (2010)

[3] A. Mirizzi et al. Rivista del Nuovo Cimento Vol. 39 N. 1-2 (2016)

Author: RUMLESKIE, Janet (Laurentian University)

Presenter: RUMLESKIE, Janet (Laurentian University)

Session Classification: M2-3 Particles deep underground (PPD) | Particules profondément sous terre (PPD)

Track Classification: Particle Physics / Physique des particules (PPD)