8th International Conference on Position Sensitive Detectors



Contribution ID: 122

Type: Poster

Electrical properties of the sensitive side in Si edgeless strip detectors developed for TOTEM Experiment

Wednesday 3 September 2008 15:10 (20 minutes)

Microstrip edgeless silicon detectors have been successfully produced and tested. These detectors are fabricated with standard planar technology, reach full sensitivity in 50 \overline m from the cut edge and can operate with high bias at room temperature. These detectors employ a newly conceived terminating structure, which, although is extremely reduced with respect to the conventional ones, still prevents the breakdown and the surface current injection at high bias. Moreover they are fully efficient up to a fluence of about 1.5x1014 p cm-2, if operated with moderate cooling. The mass production of these detectors for the TOTEM Experiment has been successfully completed. Their installation in the Roman Pots is ongoing and will allow the TOTEM Experiment to detect leading protons at 10\overline from the beam at the LHC.

Author: Dr VERBITSKAYA, ElenaPresenter: Dr VERBITSKAYA, ElenaSession Classification: Poster Session 2 - PPE & Nuclear

Track Classification: Applications in Particle Physics