



Contribution ID: 167

Type: not specified

Efficiency and time resolution of a thin gap Resistive Cylindrical Chamber

Friday 13 September 2024 12:20 (20 minutes)

The Resistive Cylindrical Chamber (RCC) is a new gaseous detector designed to overcome some limitations of Resistive Plate Chambers (RPC), broadening their application range and performance. The RCC consists of two concentric cylinders made of resistive material which define a gas gap whose thickness is almost negligible with respect to the cylinder radii. The cylinder surfaces not facing the gas volume are coated with high resistivity graphite layers, which distribute the High Voltage on the whole electrode. The signals can be read out by AC coupled pick-up strip or pads which can be positioned both in the internal cavity of the smaller cylinder and/or on the external surface of the larger cylinder. This kind of structure introduces an intrinsic two-gap configuration with independent pick-up electrodes usable to implement particles tracking and improve efficiency and time resolution. Furthermore, the cylinder structure allows to pressurize the gas inside the detector up to the elastic deformation strength of the electrodes material without mechanical efforts. The gas pressurization increases the target density with a consequent intrinsic efficiency improvement, also for very thin gas-gaps. Despite of the gas gap thickness, the detector could be always designed to produce a not uniform electric field whose gradient depends on the ratio between the gas gap thickness and the cylinder radius. In this configuration the internal electric field accelerates or dumps the avalanche discharge growth according to the electrons drift direction fixed by the high voltage polarity, extending the operability respectively to very high saturated charge distributions or to low quenched eco-friendly gas mixtures for high rate capability. With this features the RCC is a detector with very high time resolution (order of 100 ps), high efficiency (about 90% for 0.3 mm gas gap filled with 95% R134a based gas mixtures), low material budget and flexible to various applications. In this presentation we will discuss the characterization of an RCC detector with 0.3 mm gas gap, operated with standard gas mixture (95% TFE - 4.7% iC4H10 - 0.3% SF₆), performed with high energy muon beam. A reference scintillator was inserted in the cylinder cavity to optimize the acceptance of the trigger system and to highlight the potential for integrating the RCC with other detection devices. Preliminary test of operation with pressurized CO₂ have been also performed highlighting some weak points of the current configuration.

Presenter: ROCCHI, Alessandro (INFN e Universita Roma Tor Vergata (IT))

Session Classification: Applied research and new ideas (part II)