24th IEEE Real Time Conference - ICISE, Quy Nhon, Vietnam



Contribution ID: 68

Type: Oral presentation

Readout system for a strip ionization chamber dedicated for proton beam profile measurement

Tuesday 23 April 2024 16:20 (20 minutes)

This work presents a 448-channel readout system integrated into a multi-strip ionization chamber for proton beam profile measurements. Miniaturized current-input ADCs inside the detector vessel provide direct digitization of charge collected on 1.5mm cathode strips. An FPGA configures ADCs dynamically and handles acquisition triggering and data processing, enabling real-time beam analysis. This compact, flexible architecture achieves sub-millimeter spatial discrimination across the 420mm x 320mm area for efficient beam parameter quantification.

Minioral

Yes

IEEE Member

No

Are you a student?

No

Author: LIN, Ye (Shanghai Advanced Research Institute, CAS)

Presenter: LIN, Ye (Shanghai Advanced Research Institute, CAS)

Session Classification: Oral presentations

Track Classification: Data Acquisition and Trigger Architectures