



Contribution ID: 1024

Type: Oral

## Transients on Arc and Converter currents in the Multicusp Cesium Surface Conversion H<sup>-</sup> Source at LANSCE

*Monday 24 June 2019 16:00 (15 minutes)*

The Multicusp Cesium Surface Conversion H<sup>-</sup> Ion Source at the Los Alamos Neutron Science Center (LANSCE) has provided beam at ~14 mA, 120 Hz, and 10% D.F. for many years of neutron science research. Recently, random high current transients were discovered in the arc current used to ionize hydrogen in the LANSCE H<sup>-</sup> ion source, and in the converter current used to convert protons to H<sup>-</sup> ions. Most have no effect, but more severe transients can cripple beam output. Hypothesized causes are related to cesiation effects, plasma potential changes, tungsten filament evaporation/sputtering, or from the arc modulator circuitry. A dedicated study was recently done on the LANSCE H<sup>-</sup> Ion source test stand to determine the cause of these transients. Current understanding indicates that the more severe transients come from a combination of cesiation effects and plasma potential changes. The status of these current transient studies on the LANSCE H<sup>-</sup> ion source will be discussed.

**Author:** KLEINJAN, David (Los Alamos National Laboratory)

**Presenter:** KLEINJAN, David (Los Alamos National Laboratory)

**Session Classification:** 3.1 Plasma, Ion, and Electron Sources I

**Track Classification:** 3.1 Plasma, Ion and Electron Sources