

100 μ PET: an ultra-high-resolution silicon-pixel-based PET scanner

Wednesday 3 July 2024 12:00 (20 minutes)

The 100 μ PET project is developing a pre-clinical medical scanner for positron-emission tomography (PET) with ultra-high-resolution molecular imaging capabilities. The scanner is composed of multiple layers of monolithic active pixel sensors (MAPS) connected to flexible printed circuits (FPC). With pixels of 150 μ m pitch and a thickness of 280 μ m + 300 μ m (MAPS + FPC), the scanner achieves unprecedented volumetric spatial resolution of 0.02 mm³, one order of magnitude better than the best current PET scanners and uniform over the scanner's field-of-view (parallax free). The MAPS and its design features will be presented, along with the pixel read-out architecture. The construction and quality control of the scanner and its multiple detection modules, prototyped with pre-production chips and FPCs, will be showcased, and the latest imaging reconstruction with simulated high-definition mouse phantoms will be presented.

Authors: PICARDI, Antonio (Universite de Geneve (CH)); FENOGLIO, Carlo Alberto; FERRERE, Didier (Universite de Geneve (CH)); CADOUX, Frank Raphael (Universite de Geneve (CH)); IACOBUCCI, Giuseppe (Universite de Geneve (CH)); SAIDI, Jihad (Universite de Geneve (CH)); PAOLOZZI, Lorenzo (Universite de Geneve (CH)); IODICE, Luca (Universite de Geneve (CH)); VICENTE BARRETO PINTO, Mateus (Universite de Geneve (CH)); Mr CARDELLA, Roberto (Universite de Geneve (CH)); Mr CAP, Sebastien (University of Geneva (CH)); GONZALEZ SEVILLA, Sergio (Universite de Geneve (CH)); BALTUS, Terry Philippe N (Universite de Geneve (CH)); KUGATHASAN, Thanushan (Universite de Geneve (CH)); FAVRE, Yannick (Universite de Geneve (CH))

Presenter: VICENTE BARRETO PINTO, Mateus (Universite de Geneve (CH))

Session Classification: Applications & Detector Systems